

Government of West Bengal  
Mahesh Bhattacharyya Homoeopathic Medical College & Hospital  
Dr. B.N. Chakraborty Sarani, Doomurjola, Howrah-711104.  
e-mail : [principalmbhmch@gmail.com](mailto:principalmbhmch@gmail.com)

Memo No.:655 /MBHMCH/CH/PRIN/ADM/22


Dated: 03.06.2022

QUOTATION NOTICE

Sealed quotation are invited from bonafied Order supplier/Dealer/Manufacturer/ Co-operative Societies having experience in Hospital supply for **supply and installation of "Semi Automated Clinical Chemistry Analyzer" in the Dept. of Pathology** of this Institution. Quotationers are also informed that Analyzer must be with following specifications :

1. Micro controller based semi-automated analyzer to perform routine biochemistry tests (endpoint, fixed time & kinetic chemistries, multi standard curve calibration).
2. Analyzer should offer minimum of 150 or more user definable programmable parameters.
3. Analyzer should have temperature selection setting parameters programmable at 20<sup>o</sup> C – 40<sup>o</sup> C in step of 1<sup>o</sup> C for flow cell.
4. Analyzer should have 6 static interference filters (not filter wheel) with wavelength selectable from 340 - 630 n.m.
5. Analyzer should have programmable aspiration volume from 100<sub>ul</sub> – 2997<sub>ul</sub>.
6. Analyzer should have user friendly software and have facility of HELP & CALIB Key.
7. Analyzer should have facility to store 5000 test results in the memory and can recall stored results with Patient ID or with date or with date & Patient ID.
8. Instrument should be CE certified and US-FDA registered.
9. The instrument manufacture should be manufacturer or reagents, controls and calibrators, and it should be preferably of INDIAN origin.
10. Analyzer should have graphic LCD and built-in full graphic printer for printing reaction curves and test results.
11. Analyzer should have full-fledged QC menu to give daily & monthly QC with Levy Jennings graph for
12. at-least two controls per test.

Quotationers are informed that GST Registration, I.T. Return, PAN & Trade license will have to submit along with the quotation paper **addressing to the Principal & Administrator, Mahesh Bhattacharyya Homoeopathic Medical College & Hospital, Dr. B.N. Chakraborty Sarani, Doomurjola, Howrah -4, within 20.06.2022.**

  
Prof.(Dr.) Madhabananda Saha  
Principal & Administrator


Mahesh Bhattacharyya Homoeopathic Medical College & Hospital

Memo No.: 655/MBHMCH/CH/PRIN/ADM/22

Dated: 03.06.22

Copy forwarded for information and publication to their respective Notice Board to:

1. The Director of Homoeopathy, Govt. of West Bengal, Swasthya Bhaban, Salt Lake City, Kol.-64.
2. The District Magistrate, Howrah.
3. The Commissioner, Howrah Municipal Corporation, Howrah.
4. The Shabadhipati, Howrah Zilla Parishad.
5. The chief Medical Officer of Health, Howrah.
6. College website.
7. Notice board.

  
Prof.(Dr.) Madhabananda Saha  
Principal & Administrator  
Mahesh Bhattacharyya Homoeopathic  
Medical College & Hospital  
Govt. of West Bengal